

SAMSVARSERKLÆRING

Certificate of Conformity

Artikkelnummer Article Number:

10170012

Sertifikat/ Certificate:

8032GZU-001 13092

Dette dokumentet har blitt utstedt etter gjennomgang av produktets/produktenes datablad, produkttester og eventuelle tekniske konstruksjonsfiler. Produktet/produktene betraktes å tilfredstille alle nevnte normer og standarder nevnt i dette dokumentet og således anses å tilfredstille alle relevante krav for produktet/produktene

This document has been issued after reviewing the product / product data sheet, product tests and any technical construction files. The product / products are considered to meet all mentioned norms and standards referred to in this document and thus deemed to meet all relevant requirements for the product /products.

Produktnavn: ELIT 1560, 200mA kontinuitet, Isolasjonsmåler 25-12

Product name: ELIT 1560, 200mA kontinuitet,

Produktbilde/ Product image**Applikasjon/**Application**Produktbeskrivelse/** Product description

Digital isolasjonsmåler med justerbar testspenning og 99 minneplasser. CAT IV 600V

Digital insulation meter with adjustable test voltage. 99 memory slots, CAT IV 600V

Normer og standarder/ Norms and standards:

EN 61010-1:2010
EN 61010-2-030:2010
EN 61010-2-033:2012
EN 61557-1:2007
EN 61557-2:2007
EN 61557-3:2007
EN 61326-1:2013

Gjerdrum

Magne Barli Teknisk sjef/Technical Director



HELLENVEGEN 9, 2022 GJERDRUM TELEFON: 63 93 88 80
WEB: WWW.ELIT.NO MAIL: FIRMAPOST@ELIT.NO

odkjent/Approved: Dato/Date: 09.09.2016

Kontrollert av/Controlled by: Tore

Galgerud